

Sample Pilot Purge Gas Filter Assembly

This sterile barrier filter assembly is attached to the Sample Pilot prior to autoclave sterilization and protects the Sample Pilot from contamination. It must be replaced prior to each autoclave cycle. This Sample Pilot Purge Gas Filter Assembly helps ensure contamination-free operation of the bioreactor system.



FEATURES

- Disposable
- Quick to remove and replace
- User friendly

SPECIFICATIONS

Product Characterization	Specification
Max Width	2.4 in (6.1 cm)
Height	10.3 in (26.2 cm)
Max Depth	2.4 in (6.1 cm)
Weight	0.11 lbs (0.05 kg)
Required Components	Sample Pilot SP100 or SP200